

# BLOOD DONOR CHAIR

**mak** HOSPITAL  
EQUIPMENT  
Made in Indonesia



Blood Donor Chair - **31306**

## General Specification

### Construction

Combined between steel sheet (thickness 1.8mm & 2.8mm & 3.8mm) and steel pipes (Ø25.4mm). Assembled with MIG welding process.

### Mattress

Made of foam & covered by imitation leather.

### Position

Back raise and leg support is adjustable by gas spring mechanism. elevation 0 to 90°.

### Accessories

Arm rest (1 pair).  
Accessories made of combined between stainless steel (8%Ni), steel sheet and covered with foam & imitation leather.

**Painting** : Epoxy powder coating finish.

**Packaging** : Knock-Down System.

**Max. Load** : 150 Kg.

### Dimension (approx.)


(L)960 mm x (W)540 mm x (H) 520 mm (on mattress platform) /1200 mm (on backraise).

### Dimension (packaging.)

(L)1850 mm x (W) 620 mm x (H) 360 mm

*Specification subject to change without prior notice*

### Dimension (mm)



# BLOOD DONOR CHAIR - 31306

Certified By TÜV : ✓ ISO 13485 : 2003    ✓ ISO : 9001:2000

**PT. MEGA ANDALAN KALASAN**  
(ENGINEERING & MANUFACTURING COMPANY)

Marketing Office :  
Jl. Gunung Sahari Raya 51/13, Jakarta 10610, Indonesia  
Phone : +62-21-4202118 Fax : +62-21-4205368,4204871  
Email : mak@chn.net.id ; Http://www.mak-techno.com